

# Hemoglobin A1c data

(Based on CAP GH2-A 2006 Survey)

Method	No.					Low	High
	Labs	Mean	S.D.	C.V.	Median	Value	Value
GH2-03	24	8.27	0.43	5.3	8.3	7.4	8.9
	23	8.03	0.46	5.8	8.0	7.1	9.3
	156	8.09	0.26	3.2	8.1	7.3	8.9
	291	7.88	0.38	4.8	7.9	6.8	9.0
	105	8.50	0.17	2.0	8.5	8.1	9.0
	15	7.98	0.44	5.5	8.0	6.9	8.6
	20	8.43	0.16	1.9	8.4	8.1	8.8
	253	8.68	0.25	2.9	8.7	8.0	9.4
	41	8.41	0.21	2.5	8.4	7.8	8.9
	489	8.11	0.26	3.2	8.1	7.4	8.9
	15	8.33	0.63	7.6	8.2	7.2	9.4
	22	8.81	0.49	5.5	8.9	7.9	9.7
	25	8.34	0.27	3.3	8.4	7.6	8.9
	250	8.74	0.33	3.7	8.7	7.8	9.6
	62	8.05	0.38	4.7	8.0	7.2	9.1
	192	8.76	0.23	2.6	8.8	8.2	9.5
195	8.61	0.21	2.5	8.6	7.8	9.8	
25	8.16	0.30	3.7	8.2	7.6	9.0	
REFERENCE METHOD *		8.40					

more than 50% of values less than 7.9 !

more than 50% of values over 8.7 !